



# Test Certificate

## Atomizers

No.: 01 / 2018

Page: 1 von 1

Delieverer Aero Pump GmbH		Manufacturer Aero Pump GmbH	
Article Nasal pump 18x3 cos transfer 140mg		Product-Code 60025801	Color-Code transparent
Customer Lindal Schweiz GmbH Müller+Krempel AG	Customer's Order No. 06248	Customer's Product Code 14.197.01	
Batch-No. P-18/007371	Date of Production 04.12.-05.12.2018	Delivery Date 07.12.2018	
Quantity 30.000 pieces	Supplier's sample -		

**1. Measuring**

Test Parameter	target	actual value
Visible Diptube length	(95,0 ± 1) mm	95,0 mm
Test Quantity	63 pieces	

**2. Visual control**

Correspond	yes
------------	-----

**3. Uniformity of dispensed dose**

Correspond	yes
------------	-----

**4. Release**


Correspond	yes
------------	-----

**5. Documentation reference**

It is herewith confirmed that the above mentioned batch complies with all the characteristics listed above; that all mentioned tests have been carefully carried out according to the corresponding tests methods.

If for the above mentioned article deviating agreements between manufacturer and client exists these are the valid basis.

All test results are documented and these information are advised on demand.

Date: 06.12.2018	Quality Assurance J. Hafke 
---------------------	-------------------------------------------------------------------------------------------------------------------