



# Test Certificate

## Atomizers

No.: 01 / 2016  
Page: 1 von 1

Delieverer Aero Pump GmbH		Manufacturer Aero Pump GmbH	
Article Nasal pump 18x3 cos transfer 140mg		Product-Code 60025801	Color-Code transparent
Customer Lindal Schweiz GmbH Müller+Krempel AG	Customer's Order No. 05788	Customer's Product Code 14.197.01	
Batch-No. P-16/000174	Date of Production 01.06.2016	Delivery Date 08.06.2016	

Quantity 25.600 pieces	Supplier's sample -
---------------------------	------------------------

### 1. Measuring

Test Parameter	target	actual value
Visible Diptube length	(95,0 ± 1) mm	94,8 mm
Test Quantity	51 pieces	

### 2. Visual control

Correspond	yes
------------	-----

### 3. Uniformity of dispensed dose


Correspond	yes
------------	-----

### 4. Release

Correspond	yes
------------	-----

### 5. Documentation reference

It is herewith confirmed that the above mentioned batch complies with all the characteristics listed above; that all mentioned tests have been carefully carried out according to the corresponding tests methods.  
If for the above mentioned article deviating agreements between manufacturer and client exists these are the valid basis.  
All test results are documented and these information are advised on demand.

Date: 07.06.2016	Quality Assurance J. Hafke 
---------------------	---