



Test Certificate

Atomizers

No.: 01 / 2021
Page: 1 von 1

| | | | |
|--|------------------------------------|--------------------------------------|---------------------------|
| Delieverer Aero Pump GmbH | | Manufacturer Aero Pump GmbH | |
| Article Nasal pump 18x3 cos transfer 140mg | | Product-Code 60025801 | Color-Code transparent |
| Customer Lindal Schweiz GmbH Müller+Krempel AG | Customer's Order No. 06471 Pos1 | Customer's Product Code 14.197.01 | |
| Batch-No. P-20/005152 | Date of Production 31.05.2021 | Delivery Date 02.06.2021 | |
| Quantity 20.000 pieces | | Supplier's sample - | |

1 Measuring

| Test Parameter | target | actual value |
|------------------------|---------------|--------------|
| Visible Diptube length | (95,0 ± 1) mm | 95,1 mm |
| Test Quantity | 42 pieces | |

2 Visual control

| | |
|------------|-----|
| Correspond | yes |
|------------|-----|

3 Uniformity of dispensed dose


| | |
|------------|-----|
| Correspond | yes |
|------------|-----|

4 Release

| | |
|------------|-----|
| Correspond | yes |
|------------|-----|

5 Documentation reference

It is herewith confirmed that the above mentioned batch complies with all the characteristics listed above; that all mentioned tests have been carefully carried out according to the corresponding tests methods.
If for the above mentioned article deviating agreements between manufacturer and client exists these are the valid basis.
All test results are documented and these information are advised on demand.

| | |
|---------------------|--|
| Date: 01.06.2021 | Quality Assurance i.V. X.Lazaridou  |
|---------------------|--|